

Member State of OIML  
Japan



OIML Certificate No  
R76/1992-JP1-06.02

## OIML CERTIFICATE OF CONFORMITY

### Issuing authority

Name: National Institute of Advanced Industrial Science and Technology  
Address: 1-3-1, Kasumigaseki, Chiyoda-ku, TOKYO, 100-8921, Japan  
Person responsible: Hiroyuki Yoshikawa

### Applicant

Name: A&D Company, Limited  
Address: 3-23-14 Higashi-ikebukuro, Toshima-ku, Tokyo 170-0013, JAPAN

### Manufacturer of the certified pattern

Name: A&D SCALES CO.,LTD.  
Address: 162-4 INSAN-NI, DEOGSAN-MYEON, JINCHEON-GUN,  
CHUNGCHEONGBUG-DO, KOREA

### Identification of the certified pattern:

Type FG.../AD-6208...  
Further characteristics see page 2

This certificate attests the conformity of the above-mentioned pattern (represented by the samples identified in the associated test report) with the requirements of the following Recommendation(s) of the International Organization of Legal Metrology (OIML):

**R76-1**, edition 1992, including amendment 1 (1994),  
for accuracy class **III**

This certificate relates only to the metrological and technical characteristics of the pattern of the instrument concerned, as covered by the relevant OIML International Recommendation(s).

This certificate does not bestow any form of legal international approval.

Member State of OIML  
Japan



OIML Certificate No  
R76/1992-JP1-06.02



The conformity was established by tests described in the associated test report  
N° 06-01-A/R76 (58 pages) and N° 06-01-B/R76 (32 pages).

The Issuing Authority

The OIML member

Dr. H. Yoshikawa



President

Feb 27 2006

Dr. Y. Miki

Feb 27 2006

Characteristics:

| Type                 | FG...                     | AD-6208... |        |
|----------------------|---------------------------|------------|--------|
| Accuracy class       | III                       | III        |        |
| Max                  | 30 kg ~ 150 kg            | 150 kg     | 200 kg |
| e                    | 10 g ~ 50 g               | 50 g       | 100 g  |
| n                    | ≤ 3000                    | ≤ 3000     | ≤ 2000 |
| Min                  | 200 g ~ 1 kg              | 1 kg       | 2 kg   |
| Tare-balancing range | ≤ 100% of Max subtractive |            |        |
| Temperature range    | -10 °C / 40 °C            |            |        |

Important note: Apart from the mention of certificate's reference number and the name of the OIML Member State in which the certificate was issued, partial quotation of the certificate or of the associated test report is not permitted, though they may be reproduced in full.